



Republic of the Philippines  
Region X  
PROVINCE OF LANAO DEL NORTE  
**LANAO NORTE PROVINCIAL HOSPITAL**  
Lower Sagadan, Baroy, Lanao del Norte  
E-mail Address: [lnph2018@gmail.com](mailto:lnph2018@gmail.com)  
Tel. No. (063) 373-6392



November 9, 2020

**ENGR. MARIETTA BORILLO**  
Chairman, Bids and Awards Committee  
Provincial Government of Lanao del Norte

Ma'am:

Greetings!

I, Dr. Ferdinand L. Bautista a Medical Officer IV of Lanao del Norte Provincial Hospital and designated as the end-user of the ultrasound machine recommends the following revision of the specifications for the purchase of Ultrasound Machine of Lanao del Norte Provincial Hospital, to wit;

Section VII. Technical Specification (Old Specification)	Revision (New Specification)
<b>LCD Display</b> <ul style="list-style-type: none"><li>High definition display monitor with WLED backlight</li></ul>	<b>LED Display</b> <ul style="list-style-type: none"><li>High definition display monitor with LED backlight</li></ul>
High contrast dynamic range of at least >1000:1 with enhanced black levels for grayscale values.	High contrast dynamic range with enhanced black levels for grayscale values.
<b>CONTROL PANEL</b> <ul style="list-style-type: none"><li>Operate on battery for adjustment on mobile exam</li></ul>	( Omit the previous specs)
<b>Transducer Management</b> <ul style="list-style-type: none"><li>Supports instantaneous switching between transducers while maintaining current depth parameters</li></ul>	<b>Transducer Management</b> <ul style="list-style-type: none"><li>Supports instantaneous switching between transducers.</li></ul>

Thank you.

Very truly yours,

  
**DR. FERDINAND L. BAUTISTA**  
MO IV LNPH